



**Sher-e-Kashmir**  
**University of Agricultural Sciences & Technology of Jammu**  
**Faculty of Veterinary Sciences & Animal Husbandry**  
**Division of Veterinary Gynaecology & Obstetrics, R.S. Pura, Jammu**

**Notice Inviting e-tender**

On behalf of SKUAST-Jammu, tender is hereby invited through e-tendering mode from the manufacturers, reputed and authorized distributors/ dealers/ suppliers for purchase of equipment as per details given hereunder:

No.	Name of the item with brief description	Qty.	Earnest Money Deposit (CDR/FDR) (in Rs.)	Cost of Tender Document (in Rs.)
1.	Table Top CO <sub>2</sub> Incubator with Gas Cylinders	01	Rs. 30000/-	Rs 1000/-

**Last Date for submission of e-tender: 23/02/2025 by 4:00PM**

**Date of opening e-tenders: 24/02/2025**

For detailed information, please visit website <https://jktenders.gov.in/> and [www.skuast.org](http://www.skuast.org)

  
Head


**Division of Veterinary Gynaecology & Obstetrics,  
FVSc & AH, SKUAST-Jammu, R.S. Pura**

**No: AUJ/FVSc/VGO/24-25/F-FF/723-728**

**Date: 07-02-2025**

**Copy to:**

- Director Research, SKUAST-J, Main Campus Chatha for kind information. This has reference to his office letter No. AUJ/DR/24-25/F-HADP/7000-7030 dated 22/10/2024.
- Dean, FVSc & AH, SKUAST-J, R.S. Pura for kind information.
- Comptroller, SKUAST-J, Main Campus Chatha for kind information.
- Estates Officer, SKUAST-J, Main Campus Chatha for kind information and with the request to upload the NIT on <http://jktenders.gov.in>.
- I/c Data Centre, SKUAST-J, Main Campus Chatha with the request to upload the NIT on the University Website.
- SVC for timely Publication of NIT in two local dailies as per Annexure-I for wide publicity.





**Sher-e-Kashmir**  
**University of Agricultural Sciences & Technology of Jammu**  
**Faculty of Veterinary Sciences & Animal Husbandry**  
**Division of Veterinary Gynaecology & Obstetrics, R.S. Pura, Jammu**

**Notice Inviting e-tender**

On behalf of SKUAST-Jammu, tender is hereby invited through e-tendering mode from the manufacturers, reputed and authorized distributors/dealers/suppliers/firms for purchase of equipment as per details given in Annexure-I:

No.	Name of the item with brief description	Qty.	Earnest Money Deposit (CDR/FDR) (in Rs.)	Cost of Tender Document (in Rs.)
1.	Table Top CO <sub>2</sub> Incubator with Gas Cylinders	01	Rs. 30000/-	Rs 1000/-

1. The Bidding documents/ NIT can be downloaded from the website <http://jktenders.gov.in> from 08-02-2025 (10:00 A.M.) to 23-02-2025 up to 4:00 P.M.
2. a) The bids shall be deposited in electronic format on the website <http://jktenders.gov.in> from 08-02-2025 (10:00 AM) to 23-02-2025 up to 4:00 P.M. The bids received will be opened on or after 24-02-2025 at 11:00 AM. Technical bids of only those bidders will be considered that fulfill the pre-eligibility conditions by the evaluation committee.  
b) The financial bids of only those bidders will be opened online who are found technically responsive by the evaluation committee.  
c) The manufacturers/ authorized distributors/dealers/suppliers/firms are advised to go through the technical specifications mentioned in the NIT.

3. **Pre-eligibility/Technical Conditions:**

Documents required for Pre-eligibility of Bid:

- i) Copy of GST certificate.
- ii) Copy of PAN/TIN Card of the concerned manufacturers/ authorized distributors/dealers/suppliers/Firms
- iii) Undertaking of not being blacklisted by any Govt. Agency/ department.
- iv) Complete address, email, phone number, bank details of the firm and a cancelled cheque.
- v) Offers should be accompanied by a DD of Rs.1000/- (non-refundable) in favour of "**Comptroller, SKUAST-Jammu**" payable at Jammu towards the Tender fee and EMD of Rs.30000/-in the shape of CDR/FDR payable to Comptroller, SKUAST-Jammu.

**Terms & Conditions/Instructions**

1. No tender without cost of tender document (demand draft) and earnest money in the shape of CDR/FDR shall be entertained. The CDR/FDR of tenderers shall be released after the purchase process is over.
2. The earnest money shall be forfeited if any manufacturers/ authorized distributors/dealers/suppliers/firms withdraws his bid or tender during the validity period or there is failure of successful bidder to furnish the equipment and performance security within specified time period after issuance of letter of acceptance.

*[Handwritten signature]*

*[Handwritten signature]*

3. Each bid must be accompanied with the proof of Earnest Money and Tender Fee at the time of bidding and the scanned copy of the same be uploaded with the technical bid while tendering, failing which tender of the defaulter shall be rejected.
4. **To participate in bidding process:-**
  - i) The bidders have to get Digital Signature Certificate (DSC) as per Information Technology Act-2000 to participate in online bidding.
  - ii) The DSC will be required for digitally signing the bid. The bidders can get above mentioned digital certificate from any approved vendors. The bidders who already possess valid DSC need not to procure new DSC.
  - iii) The bidders have to submit their bids online in electronic format with DSC. The bids cannot be uploaded without Digital signature. No proposal will be accepted in physical form.
  - iv) Bids will be opened online as per time schedule mentioned.
  - v) Before submission of online bids, bidders must ensure that scanned copies of all the necessary documents have been attached with bid.
  - vi) The department will not be responsible for delay in online submission of bids whatsoever reasons may be.
  - vii) All the required information for bid must be filled and submitted online.
  - viii) The details of cost of documents, EMD specified in the tender documents should be the same, as submitted online (scanned copies) otherwise bid will not be accepted.
5. Any other information / clarification can be had from the office of the undersigned on all working days from 9:30 A.M. to 5:00 P.M. and email id: [vgoheadplu@gmail.com](mailto:vgoheadplu@gmail.com)
6. Conditional tenders shall be rejected outrightly.
7. The envelope in which tender is submitted must be superscribed as "Procurement of **Table Top CO<sub>2</sub> Incubator with Gas Cylinders**" for which tenders have been submitted.
8. The competent authority i.e., SKUAST - Jammu reserves the right to revise or alter the specifications of the equipment before the acceptance of the tender.
9. Late, delayed and incomplete tenders, amendments and additions other than submitted online will not be accepted.
10. The successful tenderer shall be responsible for installation of the equipment and making it fully operational.
11. The tenderer shall be fully responsible for the manufacture's warranty in respect of proper design, quality and workmanship of the equipment (s) accessories etc. covered by the tender for a period of 12 months from the date of satisfactory installation/commissioning of the system.
12. The tenderers shall make provision for imparting training to our scientists/maintenance staff on operation and use of **Table Top CO<sub>2</sub> Incubator with Gas Cylinders**.
13. The tenderer should enclose a certificate with the tender stating that:
  - i) The equipment is of the latest technology.
  - ii) The equipment shall be promptly and properly serviced by them whenever desired and such service will remain available to SKUAST-J for 01 year from the date of its installation.
15. Prices quoted by the bidder should be inclusive of transit insurance, freight, installation and commissioning at destination sites. Discount, if any, should be shown separately. Additional charges, if any, on account of Excise Duty, GST, VAT, Entry Tax or any other charge / levy must be specifically quoted.

16. The offer should be valid for a period of 30 days from the specified date of allotment of the tender.
17. 100% payment shall be made on installation and commissioning of the equipment. However, the tenderer shall have to furnish a bank guarantee equal to 10% of the total cost of the equipment towards performance guarantee for one year which shall be released on successful completion of the warranty period.
18. The company must offer model fulfilling the tender specifications. The catalogue/brochure of the equipment model should be available on the firm/company website for online verification.
19. The hard copy of tender documents including original copy of DD (cost of tender document) and CDR/FDR (EMD) must reach in the office of the **Head, Division of Veterinary Gynaecology, FVSc & AH, SKUAST-Jammu, R.S. Pura, Jammu-181102 (UT of J&K)** not later than the specified time as stipulated in the NIT, if the date on which the tenders are to be received / opened is declared as a public holiday, the tenders shall be received and opened on the next working day.
20. The competent authority of SKUAST-J does not bind itself/himself to accept the lowest or any tender & reserve the right of accepting the whole or any part of the tender or portion of the quantity offered and the bidders shall supply the same at the rate quoted.
21. The payment shall not be processed unless and until the equipment has been delivered, installed / commissioned and accepted, in accordance with the conditions of the contract and to the entire satisfaction of the competent authority of SKUAST-J.
22. All questions, disputes, or difference arising under and out of, in connection with the contract shall be subject to the Courts at Jammu (J&K).
23. In case the materials supplied by the supplier/bidder are not found as per specifications mentioned in the tender notice and other certified documents, the same shall be rejected and supplier must lift the supplies at his own risk, responsibility and cost. Further, EMD of such supplier shall be forfeited.

  
Head,

**Division of Veterinary Gynaecology & Obstetrics,  
FVSc & AH, SKUAST-Jammu, R.S. Pura**

**No: AUJ/FVSc/VGO/24-25/F-FF/723-728**

**Date: 07-02-2025**

**Copy to:**

- Director Research, SKUAST-J, Main Campus Chatha for kind information. This has reference to his office letter No. AUJ/DR/24-25/F-HADP/7000-7030 dated 22/10/2024.
- Dean, FVSc & AH, SKUAST-J, R.S. Pura for kind information.
- Estates Officer, SKUAST-J, Main Campus Chatha for kind information and with the request to upload the NIT on <http://jktenders.gov.in>.
- Comptroller, SKUAST-J, Main Campus Chatha for kind information.
- I/c Data Centre, SKUAST-J, Main Campus Chatha with the request to upload the NIT on the University Website.
- SVC for timely Publication of NIT in two local dailies as per Annexure-I for wide publicity.



### Technical specifications of "Table Top CO<sub>2</sub> Incubator with Gas Cylinders"

Description	Technical Details
Assembly	<p><b>Table Top CO<sub>2</sub> Incubator with Gas Cylinders should be equipped with:</b></p> <ol style="list-style-type: none"> <li>1. Should have at least 160 L or more capacity.</li> <li>2. Temperature management ranges from ambient+4 to 50°C with control increment of 0.1°C</li> <li>3. Temperature accuracy should be less than <math>\pm 0.5^{\circ}\text{C}</math> at 37°C, Temperature stability of <math>\pm 0.2^{\circ}\text{C}</math> at 37°C.</li> <li>4. CO<sub>2</sub> gas range should be at least 0.1 – 20% with control increment of 0.1%, gas uniformity of <math>\pm 0.1\%</math> and Stability at 5% is <math>\pm 0.1\%</math>.</li> <li>5. CO<sub>2</sub> recovery rate of 6 minutes or lesser after door opening and closing event to attain 5% CO<sub>2</sub>.</li> <li>6. Intuitive touch screen display interface.</li> <li>7. Easy access USB ports for fast data export and documentation.</li> <li>8. The operational gas pressure requirement range should be 8 -20.0 psi.</li> <li>9. Should have Dry Heat Sterilization of 180°C or more should not take more than 14 hours.</li> <li>10. Should have a display for control of temperature and alarms.</li> <li>11. The Inner chamber should be formed from single stainless-steel sheet with no corners for higher capacity and ease of cleaning.</li> <li>12. Should have direct heating elements to ensure even distribution of heat.</li> <li>13. Should come with a removable humidity tray.</li> <li>14. Should be fan-less design to reduce chance of contamination, reduce noise.</li> <li>15. Should have Infra-Red (IR); NDIR type CO<sub>2</sub> sensor and should withstand at least 180°C or more during dry heat sterilization.</li> <li>16. The incubator should come with standard 4 perforated stainless-steel shelves with 4 split inner doors.</li> <li>17. Should conform to Declaration of CE certification standards with EN 61010-1, EN 50581- ROHS Compliance.</li> <li>18. CO<sub>2</sub> Cylinder with Reputed brand regulator should be supplied.</li> <li>19. Extra accessories to be supplied viz., Online 5KVA UPS with minimum 4 hours power backup; i5 or i7 Laptop/ Computer System for easy and comprehensive documentation software for monitoring instrument performance; Three (03) extra gas cylinder; environment control system.</li> </ol>
Warranty	Minimum 1 Year

  
Head,

Division of Veterinary Gynaecology & Obstetrics,  
FVSc & AH, SKUAST-Jammu, R.S. Pura

